

Intraoperative diagnosis of *Staphylococcus aureus* and coagulase-negative *Staphylococcus* using Xpert MRSA/SA SSTI assay in prosthetic joint infection

Andrea Sambri, Giovanni Pignatti, Matteo Romagnoli, Davide Donati, Maurilio Marcacci, Matteo Cadossi

Istituto Ortopedico Rizzoli, Bologna, Italy

SUMMARY

The aim of this prospective study was to evaluate the performance of the Polymerase Chain Reaction (PCR) tool Xpert MRSA/SA SSTI test (Cepheid, Sunnyvale, CA, USA) on periprosthetic samples from a cohort of patients with suspected prosthetic joint infection (PJI).

Seventy adult patients were included in this prospective study. On the basis of the preoperative evaluation, 39 patients were clinically considered to have a PJI, whereas 31 were presumed to suffer from an aseptic mobilization of the implant. Xpert MRSA/SA SSTI identified 4 out of 4 MRSA, 7 out of 7 MSSA, and 14 out of 16 methicillin resistant CoNS.

Among the 31 patients not having a PJI, the rapid PCR did not find any bacteria among those identifiable, thus demonstrating an excellent performance in terms of specificity. Statistical analysis of the analytical performance showed a high correlation ($p < 0.001$) between the result of Xpert MRSA/SA SSTI and culture. Xpert MRSA/SA SSTI assay is a novel, yet well known, rapid and accurate method for the identification of different species of staphylococci. The test can be used with peri-operative samples thus dramatically improving the diagnostic sensitivity. In addition, thanks to the very short turnaround time the use of Xpert assay can modify the clinical management of patients suffering from PJI during the ongoing operative procedure.

Received August 22, 2016

Accepted January 23, 2017

INTRODUCTION

The incidence of prosthetic joint infection (PJI) for knee and hip has been increasing in recent years as a result of the growing number joint replacements, thus creating an economic burden on the health system of most Western countries. (Kurtzu *et al.*, 2012) The diagnosis of PJI is sometimes challenging (Del Pozo and Patel, 2009; Osmon *et al.*, 2013) and although many different clinical parameters can suggest an infection, only the identification of the infecting microorganism provides the diagnosis with the highest level of certainty.

In the case of suspected infection, bacteria can be detected by different methods. The isolation of pathogens by culture of tissue samples is still the gold standard (Achermann *et al.*, 2010; Trampuz *et al.*, 2007). Culturing the biological material dislodged from the surface of the implant with the application of ultrasound-generated power or treating the implant with dithiothreitol have both been recently reported to increase the sensitivity of organism recovery (Drago *et al.*, 2012; Trampuz *et al.*, 2007). Despite improved sensitivity, culture-negative cases remain and this prevents determination of the true presence of infec-

tion. Another major problem is the delay in obtaining a result, which can vary from 2 days to 3 weeks (Schafer *et al.*, 2008). *Staphylococcus aureus* (SA) and the group of coagulase-negative staphylococci (CoNS) are the main pathogens for PJI (ECDC, 2013) and the identification of the methicillin resistance phenotype (MR) is mandatory in order to start correct antibiotic therapy as soon as possible (Trampuz and Zimmerli, 2008).

Compared to culture, real-time polymerase chain reaction (PCR) is indeed faster and theoretically more sensitive, especially when patients are treated with antibiotics before being sampled, a well-known situation that reduces the sensitivity of culture-based methods. However, the routine use of molecular techniques in the diagnosis of PJI has been limited to selected settings by costs and complexity of the workflow (Achermann *et al.*, 2010; Dubouix-Bourandy *et al.*, 2011; Kobayashi *et al.*, 2009).

The Xpert® MRSA/SA SSTI assay (Cepheid, Sunnyvale, CA, USA) performed on the GeneXpert® instrument is a new generation test based on multiplex Real Time - PCR technology that can simultaneously detect 3 targets: staphylococcal protein A (*spa*) gene, the gene for methicillin resistance (*mecA*) and the staphylococcus cassette chromosome (*SCCmec* junction region in the staphylococcal chromosome). An internal control (SPC) (*Bacillus globigii*) is also simultaneously amplified during each diagnostic procedure to detect inhibition and proper amplification. This test is rapid (56 minutes), does not require specific technical skills and has excellent sensitivity (94.4%) and specificity (100%), as reported in the literature (Wolk *et*

Key words:

Prosthetic joint infection, Rapid diagnosis, PCR, *Staphylococcus aureus*.

Corresponding author:

Andrea Sambri MD

E-mail: andrea_sambri@libero.it

al., 2009a; 2009b). Even if the Xpert MRSA/SA SSTI assay is not validated for the detection of CoNS, when only the *mecA* gene is identified there is a strong indication of the presence of a MR CoNS.

The aim of this prospective study was to evaluate the performance of the Xpert MRSA/SA SSTI test on periprosthetic samples collected by E-swab (Copan, Brescia Italy) obtained from a cohort of patients with suspected chronic PJI and to compare the results of the molecular method with those obtained by the standard diagnostic procedure based on cultures of periprosthetic tissue samples and sonication fluids from the surface of the devices.

MATERIAL AND METHODS

Between February and November 2015, 70 adult (>18 yo) patients who underwent removal of at least one component of hip or knee arthroplasty due to septic or aseptic loosening were included in this prospective study. Patients with rupture of the prosthesis or affected by a periprosthetic fracture were excluded. The study was approved by the local ethics committee (Istituto Ortopedico Rizzoli, Bologna, Italy).

Preoperatively, all information on the location and the type of prosthesis, the clinical presentation, laboratory tests (ESR and CRP, leukocyte parameters, fibrinogen) and nuclear diagnostics (scintigraphy with labeled leukocytes) were recorded for each patient enrolled. According to the clinical presentation (presence of a fistula communicating with the implant), patients were considered to have a septic mobilization. In the absence of fistula and in the presence of elevated ESR >30 mm/h and CRP >1 mg/dl, a joint aspiration was performed for examination and patients were considered to have suspected septic or aseptic failure based on the aspirate leukocyte count (leukocytes >3000 per microliter, with >80% neutrophils).

During the surgery, at least 5 separate biopsies from periprosthetic tissue were collected. All prosthetic components were removed in the operating room under sterile conditions and transported to the microbiology laboratory where they were processed with ultrasound treatment. All the samples were collected intra-operatively before any empiric antibiotic therapy was started.

The fluids obtained by sonication and the periprosthetic samples were cultured using beadmill processing to isolate all the microorganisms present (Roux *et al.*, 2011; Trampuz *et al.*, 2007) Both aerobic and anaerobic cultures at 7 and 14 days and sensitivity test to antimicrobial drugs were performed (Molina-Manso *et al.*, 2013).

One Eswab from the periprosthetic tissue from each patient was collected and placed in the Eswab Amies transport medium in the Eswab vial before the prosthesis was

removed. Each Eswab was placed into a sampler reagent vial from the Xpert kit, vortexed for 10 seconds, and then transferred into an Xpert® MRSA/SA SSTI cartridge directly in the operating theatre.

Patients were diagnosed to have a PJI when at least one of the following criteria was present:

- a) a sinus tract communicating with the prosthesis;
- b) the same microorganism isolated in at least 2 samples (at least in 1 sample with bacterium not belonging to the cutaneous flora, such as SA);
- c) at least four of the following six elements:
 - 1) acute inflammation on the intra-operative histological examination (at least 5 neutrophils per field at high magnification);
 - 2) presence of pus in the joint;
 - 3) microorganism isolated in only one of five biopsies;
 - 4) increased white blood cell count of aspirated joint fluid >3000 per microliter;
 - 5) increased percentage of neutrophils in aspirated joint fluid >80%; 6.ESR >30 mm/h and CRP >1 mg/dl (Parvizi *et al.*, 2011).

Differences between the results obtained by culture-based techniques and those of PCR were assessed by the log-rank test. P values <0.05 were considered significant. All analyses were completed using the Statistical Package for Social Science (IBM Corp. Released 2013. IBM SPSS Statistics for Windows, Version 22.0. Armonk, NY, USA).

RESULTS

Seventy patients were included in this prospective study (39 males and 31 females). The median age was 70 years (range, 18-88); 42 (60%) patients had a hip prosthesis, 28 (40%) a knee prosthesis. On the basis of the preoperative evaluation, 39 (56%) of these patients were clinically considered to have a PJI, whereas 31 (44%) were presumed to suffer from an aseptic mobilization of the implant.

Thirty-seven patients considered to have a septic mobilization of the implant on clinical evaluation were confirmed to have a PJI according to previously described criteria (presence of fistula, blood and joint tests and cultures of sonication fluid and biopsy samples) (Table 1).

In 9 cases, the culture-based diagnosis demonstrated the presence of bacteria (MS CoNS in 3 cases, *Streptococcus agalactiae* in 2 cases and *Streptococcus gordonii*, *E. aerogenes*, *E. faecalis*, and *Pseudomonas aeruginosa* in 1 case each): all these species are not detectable by the Xpert MRSA/SA SSTI assay and were therefore excluded from the final evaluation of the performance of this test. In one additional patient neither cultural examination nor PCR test identified the pathogen responsible but PJI was confirmed by the presence of a fistula communicating with

Table 1 - Performance of Xpert MRSA/SA SSTI.

| | Culture | | | | |
|--------------------|----------|------|---------|-----------------|---------|
| | MRSA | MSSA | MR CoNS | Other Bacteria* | Sterile |
| Xpert MRSA/SA SSTI | MRSA | 4 | - | - | - |
| | MSSA | - | 7 | - | - |
| | MR CoNS | - | - | 14 | - |
| | Negative | - | - | 2 | 9 |

*MS CoNS in 3 cases, *Streptococcus agalactiae* in 2 cases and *Streptococcus gordonii*, *E. aerogenes*, *E. faecalis*, and *Pseudomonas aeruginosa* in 1 case each
 **1 case was confirmed to be PJI due to the presence of a fistula.

the implant. All these patients but one with MR *Staphylococcus epidermidis* responsible for PJI had been considered septic at pre-operative evaluation.

Xpert MRSA/SA SSTI identified 4 out of 4 MRSA and 7 out of 7 MSSA. Patients with SA infection had a mean of 3.8 cultures positive for the same bacteria (range, 2-5). Those patients with MSSA PJI had antibiotic therapy changed directly in the OR, thus reducing the empiric use of daptomycin of a mean of 6.3 days (range 5-8).

In addition, MR CoNS were extrapolated by the amplification of *mecA* gene only with the final identification of 14 out of 16 bacteria which were present on cultures (mean 3.6, range 3-5). In 2 cases the rapid PCR did not identify the pathogen responsible for PJI (MR CoNS identified on cultural examination performed on 4 and 5 samples respectively). Both these patients had laboratory values suggestive for PJI (ESR >50 mm/h and CRP >2.5 mg/dl). Among 31 patients not having a PJI, rapid PCR did not find any *Staphylococcus*, yielding excellent performance in terms of specificity.

Statistical analysis of the performance did not show any difference between Xpert MRSA/SA SSTI PCR and culture performance, but a high correlation ($P < 0.001$).

DISCUSSION

In the present series we observed a higher sensitivity of the culture-based diagnosis compared with data reported in the literature and this can be explained by our combination of cultures performed on periprosthetic tissue samples and on the fluid derived from sonication of the prosthesis in order to define PJI with certainty (Parvizi et al., 2011; Portillo et al., 2014).

Staphylococcus aureus and CoNS were found to be respon-

sible for the majority of PJI, as previously reported in the literature (ECDC, 2013; Lewis et al., 2015). However, in the present series 9 (24%) infections were attributable to other bacterial species and these were obviously missed by the PCR-based method that is limited to the detection of staphylococci.

When considering exclusively the patient population with confirmed infection by SA, the Xpert assay showed a high sensitivity and specificity, as previously reported in other studies that included also septic arthritis and spondylodiscitis patients (Dubouix-Bourandy et al., 2011; Lourtet-Hascoett et al., 2015; Titecat et al., 2012; Valour et al., 2014) (Table 2). Our results highlighted the performance of the Xpert MRSA/SA SSTI assay for the rapid identification of SA (including *mecA*) directly from the periprosthetic samples obtained and processed directly in the operating theatre. Furthermore, in the present study we also observed good performance of this molecular test for the identification of *mecA* bearing CoNS, even if slightly lower accuracy than that observed for MRSA and MSSA.

Empirical treatment covering most pathogens causing PJI is generally started soon after surgery in order to prevent a negative outcome and it is continued until an antimicrobial sensitivity test (AST) result becomes available. Since staphylococci are the most frequently bacteria isolated in PJI, recommendations for empirical treatment include agents that have high activity against these species, such as daptomycin, which is associated with various side effects and must be administered at doses higher than 6 mg/kg for long periods after the implant has been removed (Byren et al., 2012; Corona Perez-Cardona et al., 2012; Sousa et al., 2010) (Figure 1). Therefore an Xpert MRSA/SA SSTI positive result allows clinicians to rapidly adapt the antimicrobial therapy with penicillin active against staphylococci for MSSA or to appropriately maintain dap-

Table 2 - Xpert performance in the present study and review of the literature.

| | Study design | No. of patients | No. of GeneXpert positive for MRSA/PJI confirmed | GeneXpert performance for MRSA detection | No. of GeneXpert positive for MSSA/PJI confirmed | GeneXpert performance for MSSA detection | No. of GeneXpert positive for MR CoNS/PJI confirmed | GeneXpert performance for MR CoNS detection |
|--------------------------------|---------------|-----------------|--|--|--|--|---|---|
| Present study | Prospective | 70 | | | 11/11 | Sensitivity 100% Specificity 100% | 14/16 | Sensitivity 87% Specificity 100% |
| Valour et al. (2014) | Retrospective | 76 | | | 68/72 | Sensitivity 94,4% Specificity 100% | | |
| Titecat et al. (2012) | Prospective | 30 | | | 37/37 | Sensitivity 100% Specificity 91,2% | 12/13 | Sensitivity 92% |
| Dubouix-Bourandy et al. (2011) | Retrospective | 105 | 2/2 | Sensitivity 100% Specificity 100% | 16/16 | Sensitivity 100% Specificity 98,3% | 23/23 | Sensitivity 100% Specificity 95,3% |
| Lourtet-Hascoett et al. (2015) | Retrospective | 62 | | | | | 9/16 | Sensitivity 56% |




Figure 1 - Algorithm tree showing possible advantages of the use of Xpert MRSA/SA SSTI assay for a more rapid onset of target antibiotic therapy.

tomycin for MRSA, therefore leading to reduced cost of overall patient management, decreased length of hospital stay, and diminished rate of potential adverse events (Gray *et al.*, 2012; Muszbek *et al.*, 2013). Last but not least, more appropriate use of pathogen-driven therapy could contribute to the reduction of emergence of resistant bacteria (Parta *et al.*, 2010). Furthermore, higher failure rates have been seen in infections by MRSA and the early identification of high-risk patients may help in determining more adequate treatment (Bradbury *et al.*, 2009).

Conventional PCR tests have some limitations that hamper the use of these methods in the routine diagnosis of PJIs. First of all, nearly all PCR assays involve complex and time-consuming analytical processing, making them impractical for routine rapid clinical use (Achermann 2010; Rosey *et al.*, 2007; Yang *et al.*, 2008). Secondly, the use of broad-range 16S rRNA gene PCR can detect previously unknown organisms but has lower sensitivity than specific PCR and fails to detect antimicrobial resistance (Hombach *et al.*, 2010). In comparison, the GeneXpert system delivers reports directly from unprocessed samples, requires only reduced technical time of non-specialized personnel, and requires no specific separated technical area. In fact, in the present study the test was performed directly in the operating room by trained nursing personnel, with a response within less than 1 hour after the test was started.

Another potential practical use of the Xpert MRSA/SA SSTI assay should be considered. In the case of revision arthroplasty, it is of paramount importance to have as many observations as possible to exclude the presence of bacteria at the time of re-implantation (Chen *et al.*, 2014). Due to its high sensitivity and specificity both for SA and MR CoNS, the Xpert assay might be used at the time of re-implantation in the 2-stage procedure, in association with intra-operative histopathological examination in cases in which SA or MR CoNS had been previously isolated from cultures at the time of the removal of the prosthesis. The development of automated multiplex PCR assays detecting a larger panel of microorganisms commonly causing PJI such as *mecA* negative CoNS and *Propionibacterium acnes* might extend the use of Xpert also in the case of the 1-stage revision arthroplasty (Kunutsor *et al.*, 2016).

Until such a test is available, conventional cultures remain the reference method for the diagnosis of non-staphylococcal and/or plurimicrobial PJI.

We conclude that the Xpert MRSA/SA SSTI assay is a novel, yet well known, rapid, and accurate method for the determination of different species of staphylococci and their genetic resistance determinants to methicillin in peri-operative samples that can dramatically improve the diagnosis and the clinical management of patients suffering from bone and joint infections, as it can detect the most common species, as well as resistance to methicillin, during the ongoing operative procedure.

Acknowledgements

We thank Cepheid for providing test kits for this study. Cepheid had no role in the study design, data collection, data analysis or data interpretation. This research did not receive any specific grant from funding agencies in the public, commercial, or not-for-profit sectors.

References

Achermann Y, Vogt M., Leunig M., Wust J., Trampuz A. (2010). Improved diagnosis of periprosthetic joint infection by multiplex PCR of sonication fluid from removed implants. *Journal of clinical microbiology*. **48**, 1208-1214.

Bradbury T, Fehring T.K., Taunton M., Hanssen A., Azzam K., et al. (2009). The fate of acute methicillin-resistant *Staphylococcus aureus* periprosthetic knee infections treated by open debridement and retention of components. *The Journal of arthroplasty*. **24** (Suppl. 6): 101-104.

Byren I., Rege S., Campanaro E., Yankelev S., Anastasiou D., et al. (2012). Randomized controlled trial of the safety and efficacy of Daptomycin versus standard-of-care therapy for management of patients with osteomyelitis associated with prosthetic devices undergoing two-stage revision arthroplasty. *Antimicrobial agents and chemotherapy*. **56**, 5626-5632.

Chen AF, Heller S, Parvizi J. (2014) Prosthetic joint infections. *The Surgical clinics of North America*. **94**, 1265-1281.

Corona Perez-Cardona P.S., Barro Ojeda V., Rodriguez Pardo D., Pigrau Serrallach C., Guerra Farfan E., et al. (2012). Clinical experience with daptomycin for the treatment of patients with knee and hip periprosthetic joint infections. *The Journal of antimicrobial chemotherapy*. **67**, 1749-1754.

Del Pozo J.L., Patel R. (2009). Clinical practice. Infection associated with prosthetic joints. *The New England journal of medicine*. **361**, 787-794.

Drago L., Romano C.L., Mattina R., Signori V., De Vecchi E. (2012). Does dithiothreitol improve bacterial detection from infected prostheses? A pilot study. *Clinical orthopaedics and related research*. **470**, 2915-2925.

Dubouix-Bourandy A., de Ladoucette A., Pietri V., Mehdi N., Benzaquen D.,

- et al. (2011). Direct detection of Staphylococcus osteoarticular infections by use of Xpert MRSA/SA SSTI real-time PCR. *Journal of clinical microbiology*. **49**, 4225-4230.
- ECDC. European Centre for Disease Prevention and Control. Surveillance of surgical site infections in Europe 2010-2011. Stockholm 2013.
- Gray A., Dryden M., Charos A. (2012). Antibiotic management and early discharge from hospital: an economic analysis. *The Journal of antimicrobial chemotherapy*. **67**, 2297-2302.
- Hombach M., Pfyffer G.E., Roos M., Lucke K. (2010). Detection of methicillin-resistant Staphylococcus aureus (MRSA) in specimens from various body sites: performance characteristics of the BD GeneOhm MRSA assay, the Xpert MRSA assay, and broth-enriched culture in an area with a low prevalence of MRSA infections. *Journal of clinical microbiology*. **48**, 3882-3887.
- Kobayashi N., Inaba Y., Choe H., Iwamoto N., Ishida T., et al. (2009). Rapid and sensitive detection of methicillin-resistant Staphylococcus periprosthetic infections using real-time polymerase chain reaction. *Diagnostic microbiology and infectious disease*. **64**, 172-176.
- Kunutsor S.K., Whitehouse M.R., Lenguerrand E., Blom A.W., Beswick A.D., et al. (2016). Re-Infection Outcomes Following One- And Two-Stage Surgical Revision of Infected Knee Prosthesis: A Systematic Review and Meta-Analysis. *PLoS One*. **11**, e0151537.
- Kurtz S.M., Lau E., Watson H., Schmier J.K., Parvizi J. (2012). Economic burden of periprosthetic joint infection in the United States. *The Journal of arthroplasty*. **27** (Suppl. 8): 61-65 e1.
- Lewis S.S., Dicks K.V., Chen L.F., Bolognesi M.P., Anderson D.J., et al. (2015). Delay in diagnosis of invasive surgical site infections following knee arthroplasty versus hip arthroplasty. *Clinical infectious diseases: an official publication of the Infectious Diseases Society of America*. **60**, 990-996.
- Lourtet-Hascoett J., Bicart-See A., Felice M.P., Giordano G., Bonnet E. (2015). Is Xpert MRSA/SA SSTI real-time PCR a reliable tool for fast detection of methicillin-resistant coagulase-negative staphylococci in periprosthetic joint infections? *Diagnostic microbiology and infectious disease*. **83**, 59-62.
- Molina-Manso D., del Prado G., Ortiz-Perez A., Manrubia-Cobo M., Gomez-Barrena E., et al. (2013). *In vitro* susceptibility to antibiotics of staphylococci in biofilms isolated from orthopaedic infections. *International journal of antimicrobial agents*. **41**, 521-523.
- Muszbec N., Chapman R., Browne C., Marsh K., Gould I.M., et al. (2013). Using daptomycin in hospitalised patients with cSSTI caused by Staphylococcus aureus has an impact on costs. *Chemotherapy*. **59**, 427-434.
- Osmon D.R., Berbari E.F., Berendt A.R., Lew D., Zimmerli W., et al. (2013). Diagnosis and management of prosthetic joint infection: clinical practice guidelines by the Infectious Diseases Society of America. *Clinical infectious diseases: an official publication of the Infectious Diseases Society of America*. **56**, e1-e25.
- Parta M., Goebel M., Thomas J., Matloobi M., Stager C, et al. (2010). Impact of an assay that enables rapid determination of Staphylococcus species and their drug susceptibility on the treatment of patients with positive blood culture results. *Infection control and hospital epidemiology*. **31**, 1043-1048.
- Parvizi J., Zmistowski B., Berbari E.F., Bauer T.W., Springer B.D., et al. (2011). New definition for periprosthetic joint infection: from the Workgroup of the Musculoskeletal Infection Society. *Clinical orthopaedics and related research*. **469**, 2992-2994.
- Portillo M.E., Salvado M., Alier A., Martinez S., Sorli L., et al. (2014). Advantages of sonication fluid culture for the diagnosis of prosthetic joint infection. *The Journal of infection*. **69**, 35-41.
- Rosey A.L., Abachin E., Quesnes G., Cadilhac C., Pejin Z., et al. (2007). Development of a broad-range 16S rDNA real-time PCR for the diagnosis of septic arthritis in children. *Journal of microbiological methods*. **68**, 88-93.
- Roux A.L., Sivadon-Tardy V., Bauer T., Lortat-Jacob A., Herrmann J.L., et al. (2011). Diagnosis of prosthetic joint infection by beadmill processing of a periprosthetic specimen. *Clinical microbiology and infection: the official publication of the European Society of Clinical Microbiology and Infectious Diseases*. **17**, 447-450.
- Schafer P., Fink B., Sandow D., Margull A., Berger I., et al. (2008). Prolonged bacterial culture to identify late periprosthetic joint infection: a promising strategy. *Clinical infectious diseases: an official publication of the Infectious Diseases Society of America*. **47**, 1403-1409.
- Sousa R., Pereira A., Massada M., da Silva MV, Lemos R., et al. (2010). Empirical antibiotic therapy in prosthetic joint infections. *Acta orthopaedica Belgica*. **76**, 254-259.
- Titecat M., Loiez C., Senneville E., Wallet F., Dezeque H., et al. (2012). Evaluation of rapid mecA gene detection versus standard culture in staphylococcal chronic prosthetic joint infections. *Diagnostic microbiology and infectious disease*. **73**, 318-321.
- Trampuz A., Piper K.E., Jacobson M.J., Hanssen A.D., Unni K.K., et al. (2007). Sonication of removed hip and knee prostheses for diagnosis of infection. *The New England journal of medicine*. **357**, 654-663.
- Trampuz A., Zimmerli W. (2008). Diagnosis and treatment of implant-associated septic arthritis and osteomyelitis. *Current infectious disease reports*. **10**, 394-403.
- Valour F., Blanc-Pattin V., Freydiere A.M., Bouaziz A., Chanard E., et al. (2014). Rapid detection of Staphylococcus aureus and methicillin resistance in bone and joint infection samples: evaluation of the GeneXpert MRSA/SA SSTI assay. *Diagnostic microbiology and infectious disease*. **78**, 313-315.
- Wolk D.M., Picton E., Johnson D., Davis T., Pancholi P., et al. (2009a). Multicenter evaluation of the Cepheid Xpert methicillin-resistant Staphylococcus aureus (MRSA) test as a rapid screening method for detection of MRSA in nares. *Journal of clinical microbiology*. **47**, 758-764.
- Wolk D.M., Struelens M.J., Pancholi P., Davis T., Della-Latta P., et al. (2009b). Rapid detection of Staphylococcus aureus and methicillin-resistant S. aureus (MRSA) in wound specimens and blood cultures: multicenter preclinical evaluation of the Cepheid Xpert MRSA/SA skin and soft tissue and blood culture assays. *Journal of clinical microbiology*. **47**, 823-826.
- Yang S., Ramachandran P., Hardick A., Hsieh Y.H., Quianzon C., et al. (2008). Rapid PCR-based diagnosis of septic arthritis by early Gram-type classification and pathogen identification. *Journal of clinical microbiology*. **6**, 1386-1390.